

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Page 11 of 11

Proficiency Test Report

Lab Id: 11642	USGS NATIONAL WATER QUALITY LABORATORY POB 25046-MS 407-BLDG 95	Shipment Date : 24-Jan-2006
EPA Lab Code: CO00009	DENVER, CO 80225-0046 (303) 236-2000	Closing Date : 09-Mar-2006
	Director: MR. MERLE SHOCKEY	Score Date : 29-Mar-2006

This report may contain data that are not covered by the NVLAP accreditation.

** Indicates NVLAP accredited analyte evaluated using the USEPA's National Standards for Water Proficiency Testing Studies Criteria Document.
NVLAP Lab Code 200387-0.

Shipment: 290 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Method	Result	Mean/Target *	Warning Limits	Acceptance Limits	Score
Sample: Non Potable Water Purgeables								
trans-1,3-Dichloropropene EPA Code: N/A	ug/L	9013	USGS O-4127-96	< 0.9	< 8.00 *			Satisfactory <i>102 passed out of 107 reported results.</i>
Tetrachloroethene ** EPA Code: N/A	ug/L	9013	USGS O-4127-96	85.6	89.0 *	61.4 - 105	50.5 - 116	Satisfactory <i>108 passed out of 111 reported results.</i>
Toluene ** EPA Code: N/A	ug/L	9013	USGS O-4127-96	55.5	56.1 *	44.2 - 65.1	38.9 - 70.4	Satisfactory <i>120 passed out of 123 reported results.</i>
Trichloroethene ** EPA Code: N/A	ug/L	9013	USGS O-4127-96	84.0	82.2 *	61.4 - 97.8	52.3 - 107	Satisfactory <i>111 passed out of 113 reported results.</i>
Total Xylenes EPA Code: N/A	ug/L	9013	USGS O-4127-96	122	114		68.1 - 161	Satisfactory <i>113 passed out of 119 reported results.</i>
Vinyl chloride EPA Code: N/A	ug/L	9013	USGS O-4127-96	54.6	56.8 *		22.7 - 90.9	Satisfactory <i>103 passed out of 107 reported results.</i>
Sample: Non Potable Water Residue								
Solids, Total EPA Code: N/A	mg/L	9002	USGS I-3750-85	357	368		327 - 409	Satisfactory <i>99 passed out of 104 reported results.</i>
Sample: Non Potable Water Total Dissolved Solids								
Solids, Total Dissolved ** EPA Code: N/A	mg/L	9040	USGS I-1750-85	498	503 *	418 - 572	380 - 610	Satisfactory <i>105 passed out of 107 reported results.</i>